

Odhad kombinované nejistoty:

Platnost do:

28.8.2020

č.	Identifikace postupu/metody	analyzátor	Odhad kombinované nejistoty	
			k=1	k=2
1	SOPV- Albumin -Sérum	Abbott Architect c16000 (nový)	1,7%	3,3%
2	SOPV- ALP	Abbott Architect c16000 (nový)	4,3%	8,6%
3	SOPV- ALT	Abbott Architect c16000 (nový)	1,8%	3,5%
4	SOPV- AMS	Abbott Architect c16000 (nový)	2,9%	5,9%
5	SOPV- AST	Abbott Architect c16000 (nový)	2,1%	4,1%
6	SOPV- Bilirubin celkový	Abbott Architect c16000 (nový)	3,3%	6,7%
7	SOPV- Celkova bílkovina, sérum	Abbott Architect c16000 (nový)	3,0%	6,1%
8	SOPV- CRP	Abbott Architect c16000 (nový)	2,2%	4,5%
9	SOPV- Glukoza, sérum	Abbott Architect c16000 (nový)	5,5%	11,0%
10	SOPV- GMT	Abbott Architect c16000 (nový)	2,3%	4,6%
11	SOPV- HDL	Abbott Architect c16000 (nový)	2,9%	5,9%
12	SOPV- Cholesterol	Abbott Architect c16000 (nový)	1,8%	3,6%
13	SOPV- Cholinesteráza	Abbott Architect c16000 (nový)	2,9%	5,9%
14	SOPV- ionty Cl-sérum	Abbott Architect c16000 (nový)	1,9%	3,8%
15	SOPV- ionty K-sérum	Abbott Architect c16000 (nový)	1,2%	2,5%
16	SOPV- ionty Na-sérum	Abbott Architect c16000 (nový)	2,1%	4,2%
17	SOPV- Kreatinin sérum	Abbott Architect c16000 (nový)	3,3%	6,7%
18	SOPV- Kyselina močová, sérum	Abbott Architect c16000 (nový)	1,5%	3,1%
19	SOPV- Myoglobin	Abbott Architect c16000 (nový)	5,2%	10,4%
20	SOPV- Triglyceridy	Abbott Architect c16000 (nový)	2,0%	4,0%
21	SOPV- Urea-sérum	Abbott Architect c16000 (nový)	1,0%	1,9%
22	SOPV- Ca sérum	Abbott Architect c16000 (nový)	1,0%	2,0%
23	SOPV- CK	Abbott Architect c16000 (nový)	1,3%	2,5%
24	SOPV- Mg, sérum	Abbott Architect c16000 (nový)	4,2%	8,4%
25	SOPV- Albumin-moč	Abbott Architect c16000 (nový)	9,6%	19,2%
26	SOPV- Ca-moč	Abbott Architect c16000 (nový)	4,2%	8,5%
27	SOPV- Celkova bílkovina-moč	Abbott Architect c16000 (nový)	5,6%	11,2%
28	SOPV- ionty Cl-moč	Abbott Architect c16000 (nový)	1,6%	3,3%
29	SOPV- ionty K-moč	Abbott Architect c16000 (nový)	0,9%	1,8%
30	SOPV- ionty Na-moč	Abbott Architect c16000 (nový)	0,9%	1,9%
31	SOPV- Kreatinin-moč	Abbott Architect c16000 (nový)	6,3%	12,6%
32	SOPV- RF	Abbott Architect c16000 (st.)	3,1%	6,1%
33	SOPV- A1AT	Abbott Architect c16000 (st.)	3,5%	7,0%
34	SOPV- Albumin -Sérum	Abbott Architect c16000 (st.)	2,8%	5,7%
35	SOPV- ALP	Abbott Architect c16000 (st.)	4,9%	9,8%
36	SOPV- ALT	Abbott Architect c16000 (st.)	4,0%	8,1%
37	SOPV- AMS	Abbott Architect c16000 (st.)	2,4%	4,9%
38	SOPV- ApoA1	Abbott Architect c16000 (st.)	5,0%	10,0%
39	SOPV- ApoB	Abbott Architect c16000 (st.)	5,6%	11,2%
40	SOPV- ASLO	Abbott Architect c16000 (st.)	3,1%	6,3%
41	SOPV- AST	Abbott Architect c16000 (st.)	4,3%	8,6%
42	SOPV- Bilirubin celkový	Abbott Architect c16000 (st.)	3,2%	6,5%
43	SOPV- C3	Abbott Architect c16000 (st.)	2,8%	5,5%
44	SOPV- C4	Abbott Architect c16000 (st.)	2,9%	5,7%
45	SOPV- Ca sérum	Abbott Architect c16000 (st.)	2,9%	5,9%
46	SOPV- Celkova bílkovina, sérum	Abbott Architect c16000 (st.)	1,8%	3,7%
47	SOPV- CK	Abbott Architect c16000 (st.)	1,6%	3,3%
48	SOPV- CRP	Abbott Architect c16000 (st.)	1,6%	3,2%
49	SOPV- Fe	Abbott Architect c16000 (st.)	3,5%	6,9%
50	SOPV- Ferritin	Abbott Architect c16000 (st.)	2,1%	4,2%
51	SOPV- Glukoza, sérum	Abbott Architect c16000 (st.)	4,0%	7,9%
52	SOPV- GMT	Abbott Architect c16000 (st.)	1,0%	1,9%
53	SOPV- HDL	Abbott Architect c16000 (st.)	4,0%	8,1%
54	SOPV- Cholesterol	Abbott Architect c16000 (st.)	3,6%	7,2%
55	SOPV- Cholinesteráza	Abbott Architect c16000 (st.)	1,4%	2,7%
56	SOPV- IgA	Abbott Architect c16000 (st.)	2,6%	5,1%
57	SOPV- IgE	Abbott Architect c16000 (st.)	8,9%	17,9%
58	SOPV- IgG	Abbott Architect c16000 (st.)	1,4%	2,7%
59	SOPV- IgM	Abbott Architect c16000 (st.)	8,2%	16,3%
60	SOPV- ionty Cl-sérum	Abbott Architect c16000 (st.)	1,6%	3,3%
61	SOPV- ionty K-sérum	Abbott Architect c16000 (st.)	0,7%	1,4%
62	SOPV- ionty Na-sérum	Abbott Architect c16000 (st.)	1,2%	2,3%
63	SOPV- Kreatinin sérum	Abbott Architect c16000 (st.)	3,8%	7,6%
64	SOPV- Kyselina močová, sérum	Abbott Architect c16000 (st.)	6,6%	13,1%
65	SOPV- Mg, sérum	Abbott Architect c16000 (st.)	4,0%	8,1%

66	SOPV- Myoglobin	Abbott Architect c16000 (st.)	4,6%	9,2%
67	SOPV- Orosomukoid	Abbott Architect c16000 (st.)	3,5%	7,0%
68	SOPV- P-sérum	Abbott Architect c16000 (st.)	2,4%	4,8%
69	SOPV- Triglyceridy	Abbott Architect c16000 (st.)	1,5%	3,0%
70	SOPV- Urea-sérum	Abbott Architect c16000 (st.)	3,8%	7,6%
71	SOPV- AFP	Abbott Architect i1000	3,0%	6,0%
72	SOPV- HCG beta	Abbott Architect i1000	3,8%	7,7%
73	SOPV- hsTroponin I	Abbott Architect i1000	4,6%	9,1%
74	SOPV- total PSA	Abbott Architect i1000	12,5%	25,1%
75	SOPV- anti-HCV	Abbott Architect i2000	10,6%	21,2%
76	SOPV- aTG	Abbott Architect i2000	3,3%	6,6%
77	SOPV- aTPO	Abbott Architect i2000	7,1%	14,3%
78	SOPV- free T4	Abbott Architect i2000	4,9%	9,7%
79	SOPV- TSH	Abbott Architect i2000	3,0%	6,0%
80	SOPV- Glukóza chip	Biosen S-line 5222-05-0011	4,0%	8,0%
81	SOPV- Glukóza chip	Biosen S-line 5222-07-0016	4,7%	9,4%
82	SOPV- Krevní obraz - HCT	Cell-Dyn Ruby 1	2,6%	5,2%
83	SOPV- Krevní obraz - HGB	Cell-Dyn Ruby 1	1,8%	3,6%
84	SOPV- Krevní obraz - PLT	Cell-Dyn Ruby 1	2,9%	5,8%
85	SOPV- Krevní obraz - RBC	Cell-Dyn Ruby 1	1,4%	2,8%
86	SOPV- Krevní obraz - WBC	Cell-Dyn Ruby 1	6,5%	13,0%
87	SOPV- ABR - pCO2	Cobas B221	3,6%	7,1%
88	SOPV- ABR - pH	Cobas B221	0,1%	0,2%
89	SOPV- ABR - pO2	Cobas B221	9,3%	18,6%
90	SOPV- APTT	Sysmex CS 2500	2,1%	4,1%
91	SOPV- D-dimer	Sysmex CS 2500	5,4%	10,8%
92	SOPV- PT	Sysmex CS 2500	6,5%	13,0%
93	SOPV- Fibrinogen	Sysmex CS 2500	10,1%	20,1%
94	SOPV- HbA1c	Tosoh G8 11920809	1,8%	3,7%